CALIFORNIA AIR RESOURCES BOARD

FLOW CALIBRATION REPORT

SPECIAL PURPOSE MONITORING
NEIL ADLER

LOG NUMBER *2007 154*

FROM ROBERT RUSSELL\BRIAN SPREADBOROUGH
Program Evaluation & Standards

CALIBRATION DATE: *07/10/2007*
REPORT DATE : *07/10/2007*

IDENTIFICATION

Instrument : *AALBORG*
Position number : *1*
Property No. : 20062240
Serial No. : *20062240*
Previous Log No.: *N/A*
Bar Code No. : *20062240*
Elevation : *25.00'*
Inst. Prop. Of : *AIR MONITORING - CENTRAL*

Site Name : *MLD Standards Lab*
Site Number : *34-299*
Location : *1309 T-Street*
Sacramento, CA 95814

CALIBRATION STANDARDS	ID NUMBER
molbox	20062240

CALIBRATION RESULTS

Component	FLOW
Instrument Range	0-5 SLM
Initial Zero Setting	
Initial Span Setting	
Final Zero Setting	
Final Span Setting	
Slope	1.007
Intercept	-0.133
Correlation Coefficient	0.99989
Change From Previous Calibration (%)	N/A
Date Of Last Calibration	N/A

Calibration Equation

Calibration Expires: *07/09/2008*

FLOW 0.993 * (Net Display) + 0.132

Comments:

CALIBRATED BY: NB

CHECKED BY: R.R.